

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2023] 濮 社量标 濮 证字第 006 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of "Law on Metrology of the People's Republic of China."

建标单位 濮阳市质量技术监督检验测试中心

Possessor of the measurement standard

计量标准名称 汽车排放气体测试仪检定装置

代码 46113114

Name of measurement standard

Code

测量范围 HC: $(200\sim 3200) \times 10^{-6}$; CO: $(0.5\sim 4.8) \times 10^{-2}$;
Measuring range CO₂: $(3.6\sim 12.0) \times 10^{-2}$; O₂: $(0.5\sim 20.9) \times 10^{-2}$;
NO: $(300\sim 3000) \times 10^{-6}$

不确定度或准确度等级 或最大允许误差 HC: $U_{rel}=1\% k=2$; CO: $U_{rel}=1\% k=2$; CO₂: $U_{rel}=1\% k=2$;
O₂: $U_{rel}=1\% k=2$; NO: $U_{rel}=1\% k=2$

Uncertainty/accuracy class/
maximum permissible error

发证机关 (印章)

The issuing authority

发证日期 2023 年 10 月 17 日

Date issued

有效期至 2028 年 10 月 16 日

Date of expiry

