

社会公用计量标准证书

Certificate of Measurement Standard for Public Service

[2023] 濮 社量标 濮 证字第 002 号

根据《中华人民共和国计量法》第六条的规定，批准你单位建立的计量标准作为社会公用计量标准。

This is to certify that the measurement standard set up in your unit is authorized to be used for public service in accordance with article 6 of "Law on Metrology of the People's Republic of China."

建标单位 濮阳市质量技术监督检验测试中心

Possessor of the measurement standard

计量标准名称 机动车前照灯检测仪检定装置

代码 28117000

Name of measurement standard

Code

测量范围 发光强度：(0~60) kcd

Measuring range 光轴偏角：上 2° ~下 2° ，左 2° ~右 2°

不确定度或准确度等级 发光强度： $U_{rel}=6\%$ ， $k=2$

或最大允许误差 光轴偏角：MPE： $\pm 5'$

Uncertainty/accuracy class/

maximum permissible error

发证机关(印章)

The issuing authority

发证日期 2023 年 04 月 07 日

Date issued

有效期至 2028 年 04 月 06 日

Date of expiry

可开展检定或校准 项目名称 Verification or calibration items	测量范围 Measuring range	不确定度或准确度等级 或最大允许误差 Uncertainty/accuracy class/ maximum permissible error
机动车前照灯检测仪 ----以下空白----	发光强度：(0~60) kcd 光轴偏角：上 2° ~下 2° ， 左 2° ~右 2°	发光强度：MPE：±15% 光轴偏角：MPE：±15'
递地区或范围 Region of service	濮阳市辖区内	
计量标准考核证书号 Certificate No. (for examination of measurement standards)	[2023]豫量标濮证字第 002 号	
计量授权证书号 Certificate No. (for authorization) 注：由被授权单位填写	豫法计(2018)410010号	

发证机关提示：请在收到社会公用计量标准证书后核对证书内容，如有异议及时进行更正。